

Grass Seed Moisture Tables Perennial Rye Grass

Meter Reading	% Moisture
1	8.0
2	8.2
3	8.5
4	8.8
5	9.0
6	9.3
7	9.6
8	9.8
9	10.1
10	10.3
11	10.6
12	10.9
13	11.1
14	11.4
15	11.7
16	11.9
17	12.2
18	12.5
19	12.7
20	13.0
21	13.3
22	13.5
23	13.8
24	14.1
25	14.3

Meter Reading	% Moisture
26	14.6
27	14.8
28	15.1
29	15.4
30	15.6
31	15.9
32	16.2
33	16.4
34	16.7
35	17.0
36	17.2
37	17.5
38	17.8
39	18.0
40	18.3
41	18.5
42	18.8
43	19.1
44	19.3
45	19.6
46	19.9
47	20.1
48	20.4
49	20.7
50	20.9

Dry for Storage: 12%

125g. Sample

calibrate to 0 at meter reading 87

%moisture = meter reading x 2646 + 7.7005 table is correct + or -.5 %

developed by U of M (Derek Lewis), Labtronics (Jason Diehl), and MFSA

Temperature is not compensated, results assumed correct at 20 degrees C.

Funding Provided by ARDI
Developed in Cooperation with the University of Manitoba, Labtronics and
Manitoba Forage Seed Association